

# G-6-PDH CONTROL



REF 1578-0910

Packaging: 3x2x.5 mL

## INTENDED USE

Material for internal quality control of G-6-PDH assays using Medicon reagent with Diatron Pictus® P700 and P500 analyzers. For in vitro diagnostic use only by trained laboratory professionals.

## COMPOSITION

Lyophilized, non-hemolysis human blood pool from healthy donors, sterilized, filtered, de-fibrinated.

### ⚠ WARNINGS – PRECAUTIONS

- For in vitro use only.
- Contains ≤0.1% sodium azide (NaN<sub>3</sub>). Avoid contact with skin, eyes and mucous membranes.
- Every donor used for the preparation of the material was tested and found negative for HbsAg, anti-HCV and anti-HIV 1 and 2 with FDA approved methods. As there is no known test method that can offer complete assurance that products derived from human blood will not transmit infectious agents this product should be handled as a potentially infectious material
- Any serious incident that may occur in relation to this device must be reported by the user to the manufacturer and the competent authority of the country in which the user and/or the patient is established!

### ⚠ PREPARATION

Remove the rubber stopper very carefully. As the material is packed under vacuum, sudden air influx may force material out of the vial. Reconstitute by adding exactly 0.5 ml of cold (2°-8°C) deionized or distilled water. Place the rubber stopper back on the vial, swirl gently 2-3 times and let it stand for 10 minutes. Mix gently to obtain a homogeneous mixture. Ideally use a hematology tubes mixer for 20 minutes. Avoid foaming. Add 900µL of saponin 2% (MEDICON G-6-PDH reagent 3) to every 100µL sample/quality control material. Replace cap immediately after use and store at 2-8°C. Unsuitable storage, handling, or errors during the analytical procedure may give erroneous results.

### MATERIALS REQUIRED BUT NOT PROVIDED WITH THE KIT

- G-6-PDH reagent
- Automated analyzer
- Common laboratory equipment

### ⚠ STORAGE - STABILITY

Unopened the material is stable up to the expiry date stated on the label when stored at 2-8°C. After reconstitution the material is stable for 1 week when stored tightly capped immediately after use at 2-8°C.

## TEST PROCEDURE

Refer to the user's manual of the analyzer for the quality control process.

Each laboratory should establish its own control frequency, however good laboratory practice suggests that controls be tested each date patient samples are tested and each time calibration is performed. The results obtained by any individual laboratory may vary from the given mean value but should fall within the corresponding acceptable ranges given in the enclosed table. If any trends or sudden shifts in values are detected, review all operating parameters. Each laboratory should establish guidelines for corrective actions to be taken if controls do not recover within the specified limits. Make sure the LOT on the vial is the same as on the value sheet accompanying the material.

### ⚠ WASTE DISPOSAL

This product contains sodium azide (NaN<sub>3</sub>), which forms sensitive explosive compounds with copper or lead. Flush waste pipes with water after the disposal of undiluted reagent in order to avoid azide build up in the drain. Dispose of all waste material in accordance with local guidelines. Safety data sheet is available for professional use on request.

### ASSIGNED VALUES – Lot specific

Please refer to the value sheet for the specific lot available at <https://medicondoc.com>.

## SYMBOLS



Manufacturer



In vitro diagnostic medical device



Temperature Limit



Catalogue Number



Caution



Biological Risks