



DIAGNOSTIC AUTOMATION, INC.

23961 Craftsman Road, Suite D/E/F, Calabasas, CA 91302

Tel: (818) 591-3030 Fax: (818) 591-8383

onestep@rapidtest.com

technicalsupport@rapidtest.com

www.rapidtest.com



See external label



2°C-8°C



Σ=96 tests



Cat # 5135-8

Enzyme Immunoassay for the Quantitative Determination of Almond in Food

Almond

Cat # 5135-8

Sensitivity	0.25 ng/mL
Recovery (spiked samples)	85-95%
Incubation Time	200 min

GENERAL INFORMATION

Almond (*Prunus dulcis*) belongs to the rosaceae. With 25 % the fraction of proteins in almond is high. Some of these proteins are known for being allergenic. In addition to profilin which is known to be cross-reactive to grass pollen, the almond major protein AMP has the greatest relevancy. AMP is known to be heat resistant making it stable to different production processes. For this reason almond represents an important food allergen. For almond allergic persons hidden almond allergens in food are a critical problem. Already very low amounts of almond can cause allergic reactions, which may lead to anaphylactic shock in severe cases. Because of this, almond allergic persons must strictly avoid the consumption of almonds or almond containing food. Crosscontamination, mostly in consequence of the production process is often noticed. The chocolate production process is a representative example. This explains why in many cases the existence of almond residues in foods cannot be excluded. For this reason sensitive detection systems for almond residues in foodstuffs are required.

The **Diagnostic Automation, Inc. Almond ELISA** represents a highly sensitive detection system and is particularly cap-able of the quantification of almond residues in cookies, cereals, ice cream and chocolate.

PRINCIPLE OF THE TEST

The **Diagnostic Automation, Inc. Almond** quantitative test is based on the principle of the enzyme linked immunosorbent assay. An antibody directed against almond prote-ins is bound on the surface of a microtiter plate. Almond containing samples or standards are given into the wells of the microtiter plate. After 20 minutes incubation at room temperature, the wells are washed with diluted washing solution to remove unbound material. A peroxidase conjugated second antibody directed against almond proteins is given into the wells and after 20 minutes of incubation the plate is washed again. A substrate solution is added and incubated for 20 minutes, resulting in the development of a blue colour. The colour development is inhibited by the addition of a stop solution, and the colour turns yellow. The yellow colour is measured photometrically at 450 nm. The concentration of almond is directly proportional to the colour intensity of the test sample.

PRECAUTIONS

Full compliance of the following good laboratory practices (GLP) will determine the reliability of the results:

1. Prior to beginning the assay procedure, bring all reagents to room temperature (20-25°C).
2. All reagents should be mixed by gentle inversion or swirling prior to use. Do not induce foaming.
3. Once the assay has been started, all subsequent steps should be completed without interruption and within the recommended time limits.
4. Replace caps in all the reagents immediately after use. Do not interchange vial stoppers.
5. Use a separate disposable tip for each specimen to prevent cross-contamination.
6. All specimens and standards should be run at the same time, so that all conditions of testing are the same.
7. Do not mix components from different batches.
8. Do not use reagents after expiration date.
9. Check both precision and accuracy of the laboratory equipment used during the procedure (micro pipets, ELISA reader etc.).

HEALTH AND SAFETY INSTRUCTIONS

1. Do not smoke or eat or drink or pipet by mouth in the laboratory.
2. Wear disposable gloves whenever handling patient specimens.
3. Avoid contact of substrate and stop solution with skin and mucosa (possible irritation, burn or toxicity hazard). In case of contact, rinse the affected zone with plenty of water.
4. Handling and disposal of chemical products must be done according to good laboratory practices (GLP).

REAGENTS

The kit contains reagents for 96 determinations. They have to be stored at 2-8°C. Expiry data are found on the labels of the bottles and the outer package.

1. Microtiter plate consisting of 12 strips with 8 breakable wells each, coated with anti-almond antibodies.
2. Almond Standards (0; 0.4; 1; 4; 10 ppm of almond): 5 vials with 1.0 mL each, dyed red, ready-to-use
3. Conjugate (anti-almond-peroxidase): 15 mL, dyed red, ready-to-use.
4. Substrate Solution (TMB): 15 mL, ready-to-use.
5. Stop Solution (0.5 M H₂SO₄): 15 mL, ready-to-use.

6. Extraction and sample dilution buffer (Tris): 2 x 120 mL as 10x concentrate, dyed red. Dilute 1+9 with distilled water. Stored at 4°C the diluted buffer is stable for at least one week. If during the cold storage crystals precipitate, the concentrate should be warmed up to 37°C for 15 minutes.
7. Washing Solution (PBS + Tween 20): 60 mL as 10x concentrate. Dilute 1+9 with distilled water. Stored at 4°C the diluted buffer is stable for at least 4 weeks. If during the cold storage crystals precipitate, the concentrate should be warmed up to 37°C for 15 minutes.
8. Plastic bag to store unused microtiter strips.
9. Instruction Manual.

ADDITIONAL INSTRUMENTATION AND REAGENTS (not provided)

Instrumentation

- 100 - 1000 µL micropipets
- Volumetric flask
- Analytical balance
- Mortar, mixer
- Water bath
- Centrifuge
- ELISA reader (450 nm)

Reagents

- double distilled water

SAMPLE PREPARATION

Due to high risk of cross-contamination all applied instruments like applicator, mortar, glass vials etc. have to be **cleaned thoroughly** before and after each sample. Almond proteins adhere very strongly to different surfaces. To identify possible cross-contamination caused by previous extractions it is strongly recommended to note the sequence of the extractions.

The following sample preparation should be applied for all kinds of samples:

1. To maximize homogeneity and representativeness of the sample drawing, a minimum of 5 g sample should be pulverized finely in a mortar, impact mill etc.
2. 1 g of the homogenized mixture is suspended in 20 mL of **pre-diluted** extraction buffer. Afterwards the suspension is incubated for 15 min in a preheated water bath at 60°C. To ensure good homogeneity, the samples should be shaken every two minutes.
3. The samples are centrifuged for 10 minutes at 2000 g. If it is not possible to separate the supernatant from the precipitate completely, the suspension should be filtrated if necessary.
4. 100 µL of particle-free solution are applied per well. If the results of a sample are out of the measuring range, further dilution with the **pre-diluted** extraction and sample dilution buffer is necessary. The additional dilution has to be considered when calculating the concentration.

PROCEDURE

The washing solution is supplied as 10x concentrate and has to be **diluted** 1+9 with double distilled water before use.

In any case the **ready-to-use** standards provided should be determined twofold. When samples in great quantities are determined, the standards should be pipetted once before the samples and once after the samples. For final interpretation the arithmetic mean is used for calculation.

In consideration of GLP and quality control requirements a duplicate measurement of samples is recommended.

The procedure is according to the following scheme:

1. Prepare samples as described above.
2. Pipet 100 μL **ready-to-use** standards or prepared samples in duplicate into the appropriate wells of the microtiter plate.
3. Incubate for 20 minutes at room temperature.
4. Wash the plate three times as follows: Discard the contents of the wells (dump or aspirate). Pipet 300 μL of diluted washing solution into each well. After the third repetition empty the wells again and remove residual liquid by striking the plate against a paper towel. The wash procedure is critical. Insufficient washing will result in poor precision and falsely elevated absorbencies.
5. Pipet 100 μL of conjugate (anti-almond-peroxidase) into each well.
6. Incubate for 20 minutes at room temperature.
7. Wash the plate as outlined in 4.
8. Pipet 100 μL of substrate solution into each well.
9. Allow the reaction to develop in the dark (e.g. cupboard or drawer; the chromogen is light-sensitive) for 20 minutes at room temperature.
10. Stop enzyme reaction by adding 100 μL of stop solution (0.5 M H_2SO_4) into each well. The blue colour will turn yellow upon addition.
11. After thorough mixing, measure absorbance at 450 nm (reference wavelength 620 nm), using an ELISA reader. The colour is stable for 30 minutes.

CALCULATION OF RESULTS

The ready-to-use standards are prepared for a direct determination of sample concentrations. The dilution of samples in the extraction process as described in the above stated sample preparation procedure is already considered. Additional dilution due to high sample concentration has to be accounted for.

1. Calculate the average optical density (OD 450 nm) for each set of reference standards or samples.
2. Construct a standard curve by plotting the mean optical density obtained for each reference standard against its concentration in ppm on semi-log graph paper with the optical density on the vertical (y) axis and the concentration on the horizontal (x) axis. Alternatively the evaluation can be carried out by software. In this case the 4-parameter method should be preferred.
3. Using the mean optical density value for each sample, determine the corresponding concentration of almond in ppm from the standard curve. Depending on experience and/or the availability of computer capability, other methods of data reduction may be employed.

TYPICAL STANDARD VALUES

The following table contains an example for a typical standard curve. The binding is calculated as percent of the absorption of the 10 ppm standard. These values are only an example and should not be used instead of the standard curve which has to be measured in every new test.

Almond (ppm)	(% binding of 10 ppm)
10	100
4	54
1	21
0.4	13
0	6

PERFORMANCE

Sensitivity

The limit of detection (LOD) of the **Diagnostic Automation, Inc. Almond test** is 0.2 ppm.

The limit of quantification (LOQ) of the **Diagnostic Automation, Inc. Almond test** is 0.4 ppm.

Due to the variety of sample matrices and their influence on the blank, results less than the LOQ should be treated as negative.

Cross-reactivity

For the following foods no cross-reactivity could be detected:

Milk	Sunflower seeds	Walnut
Egg	Poppy seed	Pistachio
Wheat	Sesame	Chestnut
Barley	Pine nuts	Lecithin
Rye	Cashew nut	Peach
Oats	Peanut	Plum
Rice	Hazelnut	Apricot
Corn	Pecan	Cherry
Buckwheat	Brazil nut	Cocoa
Soy	Coconut	

Precision

Intra-assay Precision	3 – 9%
Inter-assay Precision	3 - 13%
Inter-lot Precision	2 – 8%

Linearity

The serial dilution of spiked samples (cookies, cereals, ice cream and chocolate) resulted in a dilution linearity of 85% - 98%.

Recovery

Mean recovery was determined by spiking samples with different amounts of almond:

Cookies	91%
Cereals	107%
Ice cream	79%
Dark chocolate	69%

REFERENCES

1. Roux KH, et al. (2001) – Detection and stability of the major almond allergen in foods. J Agric Food Chem, 49(5):2131-6
2. Venkatachalam M, et al. (2002) – Effects of roasting, blanching, autoclaving and microwave heating on antigenicity of almond (*Prunus dulcis* L.) proteins. J Agric Food Chem, 50(12):3544-8
3. Ben Rejeb S, et al. (2005) – Multi-allergen screening immunoassay for the detection of protein markers of peanut and four tree nuts in chocolate. Food Addit Contam, 22(8):709-15
4. Acosta MR, et al. (1999) – Production and characterization of rabbit polyclonal antibodies to almond (*Prunus dulcis* L.) major storage protein. J Agric Food Chem, 47(10):4053-9
5. Tawde P, et al. (2006) – Cloning and characterization of profilin (Pru du 4), a cross-reactive almond (*Prunus dulcis*) allergen. J Allergy Clin Immun, 118(4):915-922
6. De Leon MP, et al. (2003) – Immunological analysis of allergenic cross-reactivity between peanut and tree nuts. Clin Exp Allergy, 33(9):1273-80
7. Shridar K, et al. (2001) – Electrophoretic and immunological analyses of almond (*Prunus dulcis* L.) genotypes and hybrids. J Agric Food Chem, 49(4):2043-52
8. Shridar K, et al. (2002) – Biochemical characterization of amandin, the major storage protein in almond (*Prunus dulcis* L.). J Agric Food Chem, 50(15):4333-41
9. Gaur V, et al. (2008) – Purification, identification and preliminary crystallographic studies of pru du amandin, an allergenic protein from *Prunus dulcis*. Acta Cryst Sect Struct Biol Cryst Com, 64(1):32-5
10. Lee SH, et al. (2005) – A 50 kDa maize gamma-zein has marked cross reactivity with the almond major protein. J Agric Food Chem, 53(20):7965-70

Date Adopted	Reference No.
2010-01-21	DA-Almond-2010



DIAGNOSTIC AUTOMATION, INC.
23961 Craftsman Road, Suite D/E/F, Calabasas, CA 91302
Tel: (818) 591-3030 Fax: (818) 591-8383

ISO 13485-2003



Revision Date: 10/19/2010