

LISS SOLUTION

LOW IONIC STRENGTH SOLUTION

For Potentiating Serological Reactions

REF

Cont.

B05178 1x 10 mL LISS Solution

For professional in vitro diagnostic use only.

SUMMARY

Reducing the ionic strength of a test system increases the rate of red blood cell antigen-antibody binding. Low and Messeter in 1974 showed that the use of a low ionic strength solution enhances the rate of antibody uptake in first stage of agglutination, allowing incubation times to be shortened.

TEST PRINCIPLE

When used by the recommended techniques, the solution will reduce the ionic-strength of a test system, increase the rate of red blood cell antigen-antibody binding and permits a substantial reduction in incubation time and an increase in the test sensitivity with many antibody specificities (see **LIMITATIONS**).

REAGENT

DIALAB LISS Solution is a low ionic strength solution containing glycine, sodium chloride, phosphate buffer and bovine albumin. The reagent is supplied at the optimal dilution, for use with all the recommended techniques stated below without the need for further dilution or addition. For lot reference number and expiry date see **Vial Label**.

STORAGE

Reagent vials should be stored at 2-8°C on receipt. Prolonged storage at temperatures outside this range may result in accelerated loss of reagent reactivity. This reagent has undergone transport stability studies at 37°C and -25°C.

SAMPLE COLLECTION AND PREPARATION

Blood samples drawn with or without anticoagulant may be used for antigen typing. If testing is delayed, then store specimens at 2-8°C. Specimens collected into ACD, CPD or CPDA-1 may be tested up to 35 days from the date of withdrawal. Other samples should be typed within 7 days after collection. All blood samples should be washed at least twice with PBS or Isotonic saline before being tested.

CONTROLS AND ADVICE

- It is recommended Weak Anti-D and appropriate red cells (ideally R1r and rr) be tested in parallel with each batch of tests. Tests must be considered invalid if controls do not show the expected results.
- The antiglobulin technique can only be considered valid, if all negative tests react positively with IgG sensitised red cells
- The LISS solution, red cell suspensions and test sera should be at room temperature prior to use to avoid encountering unwanted positive reactions due to "cold" antibodies.
- In the **RECOMMENDED TECHNIQUES** one drop is approximately 50 µL when using the vial dropper provided.
- The use of the reagent and the interpretation of results must be carried out by properly trained and qualified personnel in accordance with the requirements of the country where the reagents are in use.
- The user must determine the suitability of the reagent for use in other techniques.

REAGENTS AND MATERIALS REQUIRED

- Anti-human globulin, i.e. DIALAB Anti-HG polyspecific/monoclonal (REF B05181) or anti-IgG
- Coombs cells washer
- Glass test tubes (10 x 75 mm or 12 x 75 mm)
- IgG sensitised red cells
- Weak Anti-D
- PBS solution (pH 6.8 – 7.2) or Isotonic saline solution (pH 6.5 – 7.5)
- Positive (ideally R1r) and negative (rr) control red cells
- Volumetric pipettes
- Water bath or dry heat incubator equilibrated to 37°C ± 2°C

RECOMMENDED TECHNIQUES

- Prepare a 2–3% suspension of washed test red cells in PBS or Isotonic saline.
- Place in a labelled test tube: 2 volumes test serum, 1 volume test red cell suspension and 2 volumes DIALAB LISS Solution.
- Mix thoroughly and incubate at 37°C for 15 minutes.
- Wash test red cells 4 times with PBS or Isotonic saline, taking care to decant saline between washes and resuspend each cell button after each wash. Completely decant saline after last wash.
- Add 2 volumes of DIALAB Anti-HG polyspecific/monoclonal to each dry cell button.
- Mix thoroughly and centrifuge all tubes for 20 seconds at 1000 rcf for a suitable alternative time and force.
- Gently resuspend the cells and read for agglutination.

INTERPRETATION OF TEST RESULTS

- Positive:** Agglutination of test red cells constitutes a positive test result.
- Negative:** No agglutination of the test red cells constitutes a negative result.

STABILITY OF THE REACTION

- Tests should be read immediately after centrifugation. Delays may result in dissociation of antigen-antibody complexes leading to false negative or weak positive reactions.
- Caution should be exercised in the interpretation of results of tests performed at temperatures other than those recommended.

LIMITATIONS

- Red cells that have a positive DAT due to a coating of IgG cannot be typed by the indirect antiglobulin technique.
- LISS Solution cannot be used with enzyme treated red cells.
- LISS Solution cannot be used as a red cell suspending medium.
- Weakly reactive Anti-A or Anti-B may not be detected using potentiating solutions.
- Some IgM antibodies requiring room temperature incubation may not be reactive under the conditions of the recommended test procedure.
- Deviation from the recommended ration of serum, cells and LISS Solution may decrease the sensitivity of the test procedure.
- Use of saline-diluted serum, or of eluates made into substrates other than fresh human serum, will result in increased ionicity and will therefore affect the sensitivity of the test.
- False positive and false negative results may occur due to improper technique or contaminated test materials.
- Not all antigen-antibody reactions are enhanced by LISS Solution.

PERFORMANCE CHARACTERISTICS

- Prior to release, each lot of DIALAB LISS Solution has been shown to enhance many antigen-antibody reactions when used by the **RECOMMENDED TECHNIQUE**.
- The solution complies with the recommendations contained in the latest issue of the Guidelines for the UK Blood Transfusion Services.

DISCLAIMER

- The user is responsible for the performance of the reagent by any method other than those mentioned in the **RECOMMENDED TECHNIQUES**.
- Any deviations from the **RECOMMENDED TECHNIQUES** should be validated prior to use.

PRECAUTIONS

- The reagent is intended for in vitro diagnostic use only.
- If vial is cracked or leaking, discard the contents immediately.
- Do not use the reagent past the expiration date (see **Vial Label**).
- Do not use the reagent if a precipitate is present.
- Protective clothing should be worn when handling the reagent, such as disposable gloves and a laboratory coat.
- The reagent has been filtered through a 0.2 µm capsule to reduce the bioburden. Once a vial has been opened the contents should remain viable up until the expiry date as long as there is no marked turbidity, which can indicate reagent deterioration or contamination.
- The reagent contains < 0.1% sodium azide. Sodium azide may be toxic if ingested and may react with lead and copper plumbing to form explosive metal azides. On disposal flush away with large volumes of water.
- Contact with LISS together with bleach causes accelerated corrosion of base metals such as copper and iron. This should be borne in mind when considering the use of bleach for decontaminating plumbing or apparatus with metal parts, which have also been in contact with LISS
- No known tests can guarantee products derived from animal sources are free from infectious agents. Care must be taken in the use and disposal of each vial and its contents.

DISPOSAL OF REAGENTS AND DEALING WITH SPILLAGES

For information on disposal of the reagent and decontamination of a spillage site see **Material Safety Data Sheets**, available on request.

LITERATURE

- Low B., Messeter L. Antiglobulin test in low ionic strength salt solution for rapid antibody screening and crossmatching. *Vox. Sang.* 1974; **26**: 53-61.
- Moore C., Mollison P.L. Use of low ionic strength saline medium in manual tests for antibody detection. *Transfusion* 1976; **16**: 291-296.
- Wicker B., Wallas C.H. A comparison of low ionic strength saline medium with routine methods for antibody detection. *Transfusion* 1976; **16**: 469-472.
- Voak D., Downie D.M., Darnborough J., Haigh T.J., Fairham S.A. Low ionic strength media for rapid antibody detection: optimum conditions and quality control. *Med. Lab. Sci.* 1980; **37**: 107-118.
- Hiagh T.J., Fairham S.A. Advantages of low ionic strength saline (LISS) techniques in blood bank management. *Med. Lab. Sci.* 1980; **37**: 119-125.
- Dynan P.K. Evaluation of commercially available low ionic strength salt (LISS) solutions. *Med. Lab. Sci.* 1981; **38**: 13-20.
- Voak D., Downie M., Haigh T.J., Cook N. Improved antiglobulin tests to detect difficult antibodies: detection of Anti-Kell by LISS. *Med. Lab. Sci.* 1982; **39**: 363-370.
- Phillips P.K., Bebbington C. The pH, conductivity and osmolality of low ionic strength solutions used within the U.K. for the antiglobulin

- test. Transfusion Medicine 1991; 1. 155-158.
9. Guidelines for the Blood Transfusion Service in the United Kingdom. H.M.S.O. Current Edition.
 10. British Committee for Standards in Haematology, Blood Transfusion Task Force. Recommendations for evaluation, validation and implementation of new techniques for blood grouping, antibody screening and cross matching. Transfusion Medicine, 1995, 5, 145-150.

TABLE OF SYMBOLS



Batch Number



In-vitro Diagnostic



Reference Nr.



Content



Expiry Date




Store At



Manufacturer



Read Pack Insert



DIALAB Produktion und Vertrieb von chemisch – technischen
Produkten und Laborinstrumenten Gesellschaft m.b.H.
IZ NOE-Sued, Hondastrasse, Objekt M55
A – 2351 Wiener Neudorf, Austria
Phone: ++43 (0) 2236 660910-0
Fax: ++43 (0) 2236 660910-30 e-mail: office@dialab.at