

Introduction**Intended Use**

The **DRG TGF- β 2 ELISA** is an enzyme immunoassay for the quantitative *in vitro diagnostic* measurement of TGF- β 2 in serum, plasma and cell culture supernatant.

Summary and Explanation

Transforming growth factor- β (TGF- β) is a multipotent Cytokine with cell- and dose-dependent activities. This molecule is produced by a number of cells and tissue types, e.g. thrombocytes, bone tissue, placenta and kidneys. TGF- β 2 is not produced by blood platelets, contrary to TGF- β 1. This potent Cytokine modulates embryonic development, bone formation, mammary development, wound healing, hematopoiesis, cell cycle progression and the production of the extracellular matrix. TGF- β 2 – null mice were shown to exhibit perinatal mortality and a wide range of developmental defects for a single gene description which include cardiac, lung, craniofacial, limb, spinal column, eye, inner ear and urogenital defects. TGF- β 2 has been shown to be a potent growth inhibit factor in the regulation of postnatal cerebellar neurons and neuroblast proliferation. TGF- β 2 has been detected in tear fluid. TGF- β 2 levels are elevated in the vitreous of patients with proliferative diabetic retinopathy. Elevated plasma levels of TGF- β 2 have been described in patients with disseminated malignant melanoma. TGF- β 2 concentrations are furthermore elevated in Parkinson's disease in ventricular cerebrospinal fluid.

PRINCIPLE of the Test

The DRG TGF- β 2 ELISA Kit is a solid phase enzyme-linked immunosorbent assay (ELISA) based on the sandwich principle.

Prior to testing the standards and patient samples are diluted in assay buffer, acidified with HCl and then neutralized with NaOH.

Afterwards, the neutralized standards and samples are added to the antibody coated (polyclonal) microtiter wells. After the first incubation the unbound sample material is removed by washing. Then a biotinylated mouse anti TGF- β 2 antibody and the Streptavidin-HRP Enzyme complex are incubated in succession. An immuno enzyme sandwich complex is formed.

After incubation the unbound conjugate is washed off. Having added the substrate solution, the intensity of colour developed is proportional to the concentration of TGF- β 2 in the patient sample.

Warnings And Precautions

1. This kit is for *in vitro* diagnostic use only.
2. For information on hazardous substances included in the kit please refer to Material Safety Data Sheets.
3. All reagents of this test kit which contain human serum or plasma have been tested and confirmed negative for HIV I/II, HBsAg and HCV by FDA approved procedures. All reagents, however, should be treated as potential biohazards in use and for disposal.
4. Avoid contact with *Stop Solution* containing 0.5 M H₂SO₄. It may cause skin irritation and burns.
5. Never pipet by mouth and avoid contact of reagents and specimens with skin and mucous membranes.
6. Do not smoke, eat, drink or apply cosmetics in areas where specimens or kit reagents are handled.



Revised 28 Nov. 2012 rm (Vers. 11.1)



7. Wear disposable latex gloves when handling specimens and reagents. Microbial contamination of reagents or specimens may give false results.
 8. Handling should be in accordance with the procedures defined by an appropriate national biohazard safety guideline or regulation.
 9. Do not use reagents beyond expiry date as shown on the kit labels.
 10. All indicated volumes have to be performed according to the protocol. Optimal test results are only obtained when using calibrated pipettes and microtiterplate readers.
 11. Do not mix or use components from kits with different lot numbers. It is advised not to exchange wells of different plates even of the same lot. The kits may have been shipped or stored under different conditions and the binding characteristics of the plates may result slightly different.
 12. Chemicals and prepared or used reagents have to be treated as hazardous waste according the national biohazard safety guideline or regulation.
13. Safety Data Sheets for this product are available upon request directly from DRG Instruments GmbH.
The Safety Data Sheets fit the demands of: EU-Guideline 91/155 EC.

Reagents

Reagents provided:

- **Microtiterwells**, 12x8 (break apart) strips, 96 wells;
Wells coated with anti-TGF- β 2 antibody (polyclonal).
- **Standard (Stock Standard)**, 1 vial, 2 mL,
Concentrations: 1000 pg/mL
see „Preparation of Reagents“.
* contains < 0.015% BND and < 0.010% MIT as preservative.
- **Assay Buffer, 10X concentrate**, 1 vial, 10 mL,
Concentrations: 0 pg/mL
see „Preparation of Reagents“.
* contains 0.015% BND and 0.010% MIT as preservative.
- **Enzyme Conjugate**, 1 vial, 11 mL, ready to use,
Mouse anti TGF antibody conjugated to Biotin.
* contains 0.03% Proclin 300, 0.015% BND and 0.010% MIT as preservative.



Revised 28 Nov. 2012 rm (Vers. 11.1)



- **Enzyme Complex**, 1 vial, 11 mL, ready to use
Streptavidin Peroxidase
* contains 0.03% Proclin 300, 0.015% BND and 0.010% MIT as preservative.
- **Substrate Solution**, 1 vial, 14 mL, ready to use,
Tetramethylbenzidine (TMB).
- **Stop Solution**, 1 vial, 14 mL, ready to use,
contains 0.5M H₂SO₄,
Avoid contact with the stop solution. It may cause skin irritations and burns.
- **Wash Solution**, 1 vial, 30 mL (40X concentrated),
see „Preparation of Reagents“.
- 1 N **HCl**, 1 vial, 2 mL, ready to use,
for acidification of the samples.
- 1 N **NaOH**, 1 vial, 2 mL, ready to use,
for neutralization.

- * BND = 5-bromo-5-nitro-1,3-dioxane
- MIT = 2-methyl-2H-isothiazol-3-one

Note: Additional *Assay Buffer* for sample dilution is available upon request.

Material required but not provided

- 1,5 mL-Reaction Caps (e.g. from Eppendorf) for sample preparation (acidification and neutralization).
- A microtiter plate calibrated reader (450 ± 10 nm) (e.g. the DRG Instruments Microtiter Plate Reader).
- Calibrated variable precision micropipettes.
- Absorbent paper.
- Distilled or deionized water
- Universal indicator paper.
- Timer
- Semi logarithmic graph paper or software for data reduction

Storage Conditions

When stored at 2-8°C unopened reagents, except the *Stock Standard*, will retain reactivity until expiration date. Do not use reagents beyond this date.

Immediately after receipt the *Stock Standard* has to be stored frozen at -20°C.

Opened reagents must be stored at 2-8°C, or as described in chapter 4.4.

Microtiter wells must be stored at 2-8°C. Once the foil bag has been opened, care should be taken to close it tightly again. Opened kits retain activity for 6 weeks if stored as described above.

Reagent Preparation

Allow all reagents and required number of strips to reach room temperature prior to use.



Revised 28 Nov. 2012 rm (Vers. 11.1)

**Assay Buffer**

Dilute 10 mL of concentrated *Assay Buffer* with 90 mL deionized water to a final volume of 100 mL Working Assay Buffer

Standards

Serial Dilution of the *Stock Standard* (1,000 pg/mL):

<u>Description</u>		<u>Concentration</u>
Standard A		1,000 pg/mL
Standard B	1 mL Standard A + 1 mL Assay Buffer	500 pg/mL
Standard C	1 mL Standard B + 1 mL Assay Buffer	250 pg/mL
Standard D	1 mL Standard C + 1 mL Assay Buffer	125 pg/mL
Standard E	1 mL Standard D + 1 mL Assay Buffer	62.5 pg/mL
Standard F	1 mL Standard E + 1 mL Assay Buffer	31.25 pg/mL
Standard G	2 mL Assay Buffer	0 pg/mL

Note: The diluted standards are stable for 1 week at 2-8°C. For longer storage freeze at -20°C.

Wash Solution

Add deionized water to the 40X concentrated Wash Solution.

Dilute 30 mL of concentrated *Wash Solution* with 1170 mL deionized water to a final volume of 1200 mL.

The diluted Wash Solution is stable for 2 weeks at room temperature.

Disposal of the Kit

The disposal of the kit must be made according to the national regulations. Special information for this product is given in the Material Safety Data Sheets (see chapter 13).

Damaged Test Kits

In case of any severe damage to the test kit or components, DRG has to be informed in writing, at the latest, one week after receiving the kit. Severely damaged single components should not be used for a test run. They have to be stored until a final solution has been found. After this, they should be disposed according to the official regulations.



Revised 28 Nov. 2012 rm (Vers. 11.1)



SPECIMEN Collection and Preparation

Serum or plasma (EDTA- or citrate plasma) and cell culture supernatant can be used in this assay.

Do not use haemolytic, icteric or lipaemic specimens.

Please note: Samples containing sodium azide should not be used in the assay.

Specimen Collection

Serum:

Collect blood by venipuncture (e.g. Sarstedt Monovette # 02.1388.001), allow to clot, and separate serum by centrifugation at room temperature. Do not centrifuge before complete clotting has occurred. Patients receiving anticoagulant therapy may require increased clotting time.

Plasma:

Whole blood should be collected into centrifuge tubes containing anti coagulant and centrifuged immediately after collection.

(E.g. for EDTA plasma Sarstedt Monovette – red cap - # 02.166.001;

for Citrate plasma Sarstedt Monovette – green cap - # 02.167.001.)

Specimen Storage and Preparation

Specimens should be capped and may be stored for up to 24 hours at 2-8°C prior to assaying.

Specimens held for a longer time should be frozen only once at -20°C prior to assay. Thawed samples should be inverted several times prior to testing.

Specimen Dilution

SERUM AND PLASMA

Serum and Plasma Samples should be diluted **1:50** with *Assay Buffer* prior to testing.

Please note: The results have to be multiplied with the dilution factor (x 50).

Example:

dilution 1:50: 10 µL Serum + 490 µL *Assay Buffer* (mix thoroughly)

CELL CULTURE SAMPLES

Centrifuge the Cell Culture Samples. Dilute the supernatant with *Assay Buffer*, according to the expected TGF-β2 concentrations, e.g. 1:10, if a high TGF-β2 concentration is expected. The results have to be multiplied with the dilution factor.

Example:

dilution 1:10: 10 µL Sample + 90 µL *Assay Buffer* (mix thoroughly)



Revised 28 Nov. 2012 rm (Vers. 11.1)



If in an initial assay, a specimen is found to contain more than the highest standard, the specimens can be further diluted with *Assay Buffer* and reassayed as described in Assay Procedure.

For the calculation of the concentrations this additional dilution factor has to be taken into account.

Acidification and Neutralization of Samples and Standards

1. Add 200 μ L Standards or **prediluted** Sample into Reaction Caps (e.g. Eppendorf-Caps).
Please note: The standards, which have been prepared by serial dilution of the stock standard should also be prepared as described below.
2. Add 20 μ L 1 M HCl to all caps
3. Close caps, mix thoroughly (vortex) and let stand for 15 minutes
4. Add 20 μ L 1M NaOH for neutralization and mix thoroughly.
5. After neutralization the sample should have a pH value between 7 and 8. Therefore please check the pH value with Universal Indicator Paper!

Assay procedure

General Remarks

- All reagents and specimens must be allowed to come to room temperature before use. All reagents must be mixed without foaming.
- Once the test has been started, all steps should be completed without interruption.
- Use new disposal plastic pipette tips for each standard, control or sample in order to avoid cross contamination.
- Absorbance is a function of the incubation time and temperature. Before starting the assay, it is recommended that all reagents are ready, caps removed, all needed wells secured in holder, etc. This will ensure equal elapsed time for each pipetting step without interruption.
- As a general rule the enzymatic reaction is linearly proportional to time and temperature.

Test Procedure

Each run must include a standard curve.

1. Secure the desired number of Microtiter wells in the frame holder.
2. Dispense **100 µL** of each pretreated *Standard, Control* and *samples with new disposable tips* into appropriate wells.
(Please refer to chapters “Specimen Dilution” and “Acidification and Neutralization of Samples and Standards”.)
3. Cover the plate and incubate 3 hours at room temperature.
4. Briskly shake out the contents of the wells. Rinse the wells **3 times** with diluted wash solution, 300 µl per Well. Strike the wells sharply on absorbance paper to remove residual droplets.
Important note: The sensitivity and precision of this assay is markedly influenced by the correct performance of the washing procedure!
5. Dispense **100 µl** *Enzyme Conjugate* into each well.
6. Incubate **120 minutes** at room temperature.
7. Briskly shake out the contents of the wells. Rinse the wells **3 times** with diluted wash solution, 300 µl per Well. Strike the wells sharply on absorbance paper to remove residual droplets.
8. Dispense **100 µl** *Enzyme Complex* into each well.
9. Incubate **20 minutes** at room temperature.
10. Briskly shake out the contents of the wells. Rinse the wells **3 times** with diluted wash solution, 300 µl per Well. Strike the wells sharply on absorbance paper to remove residual droplets.
11. Add **100 µL** of *Substrate Solution* to each well.
12. Incubate for **10 minutes** at room temperature.
13. Stop the enzymatic reaction by adding **50 µL** of *Stop Solution* to each well.
14. Determine the absorbance (OD) of each well at **450 ± 10 nm** with a microtiter plate reader.
It is recommended that the wells be read **within 10 minutes** after adding the *Stop Solution*.

Calculation of Results

1. Calculate the average absorbance values for each set of standards, controls and patient samples.
2. Construct a standard curve by plotting the mean absorbance obtained from each standard against its concentration with absorbance value on the vertical(Y) axis and concentration on the horizontal (X) axis.
3. Using the mean absorbance value for each sample determine the corresponding concentration from the standard curve.
4. Automated method: The results in the IFU have been calculated automatically using a 4 PL (4 Parameter Logistics) curve fit. 4 Parameter Logistics is the preferred method. Other data reduction functions may give slightly different results.
5. The concentration of the samples can be read directly from this standard curve. Multiply the results by the initial dilution factor (for serum serum and plasma samples by 50)
Samples with concentrations higher than that of the highest standard have to be further diluted. For the calculation of the concentrations this dilution factor has to be taken into account.

Expected Normal values

It is strongly recommended that each laboratory should determine its own normal and abnormal values.

Data can be obtained by request.

Quality Control

Good laboratory practice requires that controls be run with each calibration curve. A statistically significant number of controls should be assayed to establish mean values and acceptable ranges to assure proper performance.

It is recommended to use control samples according to state and federal regulations. The use of control samples is advised to assure the day to day validity of results. Use controls at both normal and pathological levels.

The controls and the corresponding results of the QC-Laboratory are stated in the QC certificate added to the kit. The values and ranges stated on the QC sheet always refer to the current kit lot and should be used for direct comparison of the results.

Employ appropriate statistical methods for analysing control values and trends. If the results of the assay do not fit to the established acceptable ranges of control materials patient results should be considered invalid.

In this case, please check the following technical areas: Pipetting and timing devices; photometer, expiration dates of reagents, storage and incubation conditions, aspiration and washing methods.

After checking the above mentioned items without finding any error contact your distributor or DRG directly.

Performance Characteristics

Assay Dynamic Range

The range of the assay is between 10 – 1000 pg/mL.



Revised 28 Nov. 2012 rm (Vers. 11.1)

**Specificity of Antibodies (Cross Reactivity)**

The following substances were tested for cross reactivity of the assay:

Component	Cross reactivity
TGF-β1	none
TGF-β3	none

Sensitivity

The analytical sensitivity was calculated from the mean plus two standard deviations of twenty (20) replicate analyses of *Standard 0* and was found to be 10 pg/mL.

Reproducibility

Data can be obtained on request.

Recovery

Data can be obtained on request.

Linearity

Data can be obtained on request.

Limitations of Use

Reliable and reproducible results will be obtained when the assay procedure is performed with a complete understanding of the package insert instruction and with adherence to good laboratory practice.
Any improper handling of samples or modification of this test might influence the results.

Interfering Substances

Data can be obtained on request.

Drug Interferences

Until today no substances (drugs) are known to us, which have an influence to the measurement of TGF-β2 in a sample.

High-Dose-Hook Effect

Data can be obtained on request.

Legal Aspects**Reliability of Results**

The test must be performed exactly as per the manufacturer's instructions for use. Moreover the user must strictly adhere to the rules of GLP (Good Laboratory Practice) or other applicable national standards and/or laws. This is especially relevant for the use of control reagents. It is important to always include, within the test procedure, a sufficient number of controls for validating the accuracy and precision of the test.

The test results are valid only if all controls are within the specified ranges and if all other test parameters are also within the given assay specifications. In case of any doubt or concern please contact DRG.

**Revised 28 Nov. 2012 rm (Vers. 11.1)****Therapeutic Consequences**

Therapeutic consequences should never be based on laboratory results alone even if all test results are in agreement with the items as stated under point 11.1. Any laboratory result is only a part of the total clinical picture of a patient. Only in cases where the laboratory results are in acceptable agreement with the overall clinical picture of the patient should therapeutic consequences be derived.

The test result itself should never be the sole determinant for deriving any therapeutic consequences.

Liability

Any modification of the test kit and/or exchange or mixture of any components of different lots from one test kit to another could negatively affect the intended results and validity of the overall test. Such modification and/or exchanges invalidate any claim for replacement.

Claims submitted due to customer misinterpretation of laboratory results subject to point 11.2. are also invalid.

Regardless, in the event of any claim, the manufacturer's liability is not to exceed the value of the test kit. Any damage caused to the test kit during transportation is not subject to the liability of the manufacturer.

REFERENCES

1. Lawrence Da. Transforming growth factor-beta: An overview. *Kidney Int* 1995;47:S19-S23.
2. Roberts AB, Sporn MB. the transforming growth factor-beta. In: Sporn MB, Roberts AB., eds. *Peptide growth factors and their receptors. Handbook of experimental pharmacology*, Heidelberg: Springer Verlag, 1990:419-472.
3. Muller F, Aukrust P, Nilssen DE, Froland SS. Reduced serum level of transforming growth factor-beta in patients with IgA deficiency. *Clin Immunol Immunopathol* 1995;76:203-208.
4. Jiang X, Kanai H, Himomura K, Sawamura M, Yano S. Increased intraplatelet and urinary transforming growth factor-beta in patients with multiple myeloma. *Acta Haematol* 1995;94:1-6.
5. Taketazu F, Miyagawa K, Ichijo H, et al. Decreased level of transforming growth factor beta in blood lymphocytes of patients with aplastic anemia. *Growth Factors* 1992; 6; 85-90.
6. Snowden N, Coupes B, Herrick A, Illingworth K, Jayson MIV, Brenchley PEC. Plasma TGF beta in systemic sclerosis: A cross-sectional study. *Ann Rheum Dis* 1994;53:763-767.

Rev. 11/12/12cc