

**This kit is intended for Research Use Only.**

**Not for use in diagnostic procedures.**

### Intended Use

The Anti-Ovary Antibody ELISA Ig-Classifying test is a reliable test for measurement of immunoglobulin class specific antibodies directed against human ovaries.

This test is intended for the use with serum.

### Principles of the Assay Method

The Anti-Ovary Antibody ELISA (Enzyme Linked ImmunoSorbent Assay) Ig-classifying test is a solid-phase sandwich enzyme-immunoassay for the quantitative determination of anti-ovary antibodies in human serum.

The ELISA-plate is coated with a mix of ovary proteins which are recognized by anti-ovary antibodies. The samples and controls are pipetted into the wells and then incubated. During this incubation anti-ovary antibodies bind to the antigen and are thus immobilized on the plate. A conjugate consisting of antibodies directed against different regions of human immunoglobulins of different classes (IgA, IgG, IgM) and POD binds to the antigen-antibody-complex during the incubation. After removal of the unbound conjugate by washing the horseradish peroxidase oxidizes the then added substrate TMB (3,3',5,5'-tetramethylbenzidine) yielding a color reaction which is stopped with 0.25 M sulphuric acid (H<sub>2</sub>SO<sub>4</sub>). The extinction is measured at a wavelength of 450 nm with a microplate reader. The use of a reference measurement with a wavelength  $\geq 550$  nm is recommended, but not indispensable.

### Reagents

(sufficient for 96 determinations)

1. <b>Microtiter strips</b> coated with ovary antigen	96 wells
2. <b>Positive Control</b> , IgA, IgG, IgM	1.0 ml
3. <b>Negative Control</b> , IgA, IgG, IgM	1.0 ml
4. <b>Dilution Buffer</b> (also used as blank / zero standard / 0 U/ml )	50 ml
5. <b>Wash Solution</b> (10x concentrated)	50 ml
6. <b>Enzyme Conjugate</b> (ready for use)	
– Anti-IgG	2.5 ml
– Anti-IgA	2.5 ml
– Anti-IgM	2.5 ml
7. <b>TMB Substrate Solution</b> (solution of TMB, ready for use)	13 ml
8. <b>Stop Solution</b> (0.25 mol/l H <sub>2</sub> SO <sub>4</sub> )	13 ml

**Revised 7 April 2008 rm (Vers. 1.1)**

9. Holder for single strips 1 x

**Materials Required but not Included**

1. Microplate reader with 450 nm filter, optionally with a reference filter  $\geq 550$  nm.
2. Microliter pipettes with disposable tips: 5  $\mu$ l, 10  $\mu$ l, 50  $\mu$ l, 100  $\mu$ l, 500  $\mu$ l and 1000  $\mu$ l.
3. Tubes for the dilution of the samples
4. Distilled or demonized water
5. Absorbent paper.

Please use only calibrated pipettes and instruments.

**Warnings and Precautions**

1. This kit is intended for research use only.
2. Avoid contact with the stop solution; it may cause skin irritations and burns.
3. Do not pipette reagents by mouth.
4. Please regard all samples as potentially infectious and handle them with utmost care.
5. Handling and disposal should be in accordance with the procedures defined by an appropriate national biohazard safety guideline or regulation where this exists.

**Instructions for Reagent Preparation**

1. The components of this kit are intended for use as an integral unit and should not be interchanged with the components of other kits.
2. All reagents and specimens must be brought to room temperature before use.
3. All reagents have to be mixed without foaming.
4. Once the test procedure has been started, all steps should be continued without interruption.
5. Pipette all reagents and samples onto the bottom of the wells. Mixing or shaking after pipetting is not required.
6. Use new disposable tips for each specimen.
7. Before starting the assay, all reagents to be used should be prepared and ready for immediate use, all needed strips should be secured in the holder etc. This will ensure equal time periods for each pipetting step without interruption.
8. For optimal results it is important to wash the wells thoroughly after incubation and to remove even the last water drops by hitting the plate on absorbent paper or cloth.
9. Since the kinetics of the enzymatic reaction depends on the surrounding temperature different extinctions correlating with the respective room temperature may be observed. The optimum laboratory room temperature is 20 °C – 22 °C (68 °F – 72 °F).
10. It is recommended to effect all tests in double determination in order to minimize the consequences of pipetting or handling errors.

Revised 7 April 2008 rm (Vers. 1.1)

**Storage Instructions and Shelf Life Information**

1. Store the reagents at 2 °C – 8 °C (36 °F – 46 °F).
2. The reagents remain stable until the expiration date of the kit.
3. The diluted washing solution is stable for 4 weeks at refrigerator temperatures (4 °C – 8 °C / 39 °F – 46 °F).
4. Put caps back on the vials immediately after use.
5. Store the microtiter strips in a dry bag with desiccants. The remaining strips must be stored in the tightly resealed bag together with the desiccants. Under these storage conditions, they are stable at least for 4 weeks after opening of the sealed bag.

**Sample Material**

Serum

**Specimen Collection and Preparation**

Collect blood by venipuncture, allow to clot, and separate serum by centrifugation at room temperature; avoid haemolysis. Avoid repeated freezing and thawing. Store tubes closed as they may be a danger of contamination or alteration of concentration.

1. Handle all samples with utmost care since they may be infectious.
2. There are no known interferences with extrinsic factors or other substances.
3. Samples may be stored at different temperatures for the following time-spans:
  - Environmental temperature up to 30 °C (86 °F): up to three days
  - Refrigerator temperature (2 – 8 °C / 36 °F – 46 °F): up to one week
  - Household freezer temperature (-10 °C – -20 °C / 14 °F – -4 °F): up to one year

**ATTENTION!** There are no test methods available which may guarantee that Hepatitis B virus, Human Immunodeficiency Virus (HIV/HTLV-III/LAV), or other infectious agents are absent from the reagents in this kit. Therefore, all human blood products, including specimen samples, should be considered potentially infectious.

**Assay Procedure**

1. Warm all reagents to room temperature and mix thoroughly before use.
2. Preparation of the washing solution (10x): Dilute the concentrated washing solution (50 ml) by adding 450 ml distilled or demonized water. **Attention:** Do not use unpurified tap water!
3. Dilute sera 1: 100 with dilution buffer (1:100 dilution: 5 µl of serum + 495 µl of dilution buffer).
4. Fix the required number of coated wells or strips in the strip holder.
5. Pipette 50 µl of controls into the respective wells intended for control determination of IgA, IgM and IgG.

**Revised 7 April 2008 rm (Vers. 1.1)**

6. Pipette 50 µl of diluted serum with new disposable tips into the respective wells.
7. Incubate for 60 min at 37 °C.
8. Briskly shake out the contents of the wells and then rinse the wells 3 times with 200 µl diluted washing solution.
9. Knock the residual water out of the wells by hitting them (in the holder) on absorbent paper or cloth.
10. Dispense 50 µl of the enzyme conjugate (Anti-IgA, Anti-IgG, Anti-IgM) into each well.
11. Incubate for 60 min at 37 °C.
12. Briskly shake out the contents of the wells and then rinse the wells 5 times with 200 µl diluted washing solution.
13. Knock the residual water out of the wells by hitting them (in the holder) on absorbent paper or cloth.
14. Dispense 50 µl of substrate solution immediately after the washing to each well.
15. Incubate for 30 min at room temperature.
16. Stop the enzymatic reaction by adding 50 µl of stop solution into each well in the same sequence and time interval as dispensing the substrate.
17. Measure the extinction of the samples at 450 nm. It is recommended to carry out the measurement of the extinction within 10 minutes after stopping the reaction.

As a general rule the enzymatic reaction is linearly proportional to time and temperature. This makes interpolation possible for fixed physico-chemical conditions.

Since calibrators are assayed in each run, absorbance fluctuations do not affect the absolute results. In any case it is highly recommended to use an additional internal control if available.

**Pipetting Scheme for the Ovary Antibody ELISA Ig-Classifying Test**

	IgA				IgG				IgM			
	1	2	3	4	5	6	7	8	9	10	11	12
A	BL	BL	P	6	BL	BL	P	6	BL	BL	P	6
B	P	C	P	7	P	C	P	7	P	C	P	7
C	N	C	P	8	N	C	P	8	N	C	P	8
D	P	1	P	9	P	1	P	9	P	1	P	9
E	P	2	P	10	P	2	P	10	P	2	P	10
F	P	3	P	11	P	3	P	11	P	3	P	11
G	P	4	P	12	P	4	P	12	P	4	P	12
H	P	5	P	13	P	5	P	13	P	5	P	13

In this pipetting scheme the recommended positions for the blank (BL, please use the dilution buffer included in this kit), positive control (PC), negative control (NC), and for the specimen samples (P1 – P13) are shown as double determinations.

**Revised 7 April 2008 rm (Vers. 1.1)****Calculation of the Results**

Any microplate reader of determining the absorbency at 450 nm may be used. The determination of the reaction of each specimen sera is obtained as follows:

Two quotients (Q1/Q2) have to be formed by dividing the following optical densities:

Q1 = positive control : negative control

Q2 = specimen : negative control

The value of Q2 compared to the value of Q1 indicates whether the specimen are to be considered as positive or negative. Is Q2 higher than Q1 it is considered to be positive. The higher the value of Q2, the higher is the intensity of the positive reaction.

**Limitations of the Assay**

- At temperatures higher than 30 °C (86 °F) the samples should be transported cooled or refrigerated. The time to stop the (enzymatic color) reaction may have to be adjusted (shortened).