

APTT-P

KIT FOR DETERMINATION OF APTT (PARTICULATE ACTIVATOR)



Kit name	Kit size	Cat. No
APTT-P-4	10 x 4 ml	K-320
APTT-P-10	10 x 10 ml	K-350

INTRODUCTION

From its origins through the work of Langdell and coworkers, and later modified by others, the Activated Partial Thromboplastin Time test (APTT, kaolin-kephalin time) has been widely used for a number of years as a pre-surgical screen for assessing certain coagulation factors and in monitoring Heparin therapy. All Factors of the Intrinsic Pathway are necessary for normal results when performing the APTT test.

The APTT test is also used to monitor Heparin therapy, showing prolonged test results at approximately 0.1 units and above.

The test is also used in the quantitative determination (Factor Assays) of Factors VIII, IX, XI, XII and Prekallikrein.

METHOD PRINCIPLE

APTT-P-REAGENT contains phospholipid and magnesium aluminium silica as the activator for optimum sensitive Factor deficiencies and to heparin. The APTT-P test measures the clotting time of test plasma after the addition of APTT-P-REAGENT and followed by the addition of calcium chloride. Deficiencies of approximately 40% and lower of Factors VIII, IX, XI and XII will result in a prolonged APTT. Heparin also results in a prolonged APTT.

REAGENTS

Package

	APTT-P-4	APTT-P-10
APTT-P-REAGENT	10 x 4 ml	10 x 10 ml

Working reagent preparation and stability

The reagents are ready to use.

The reagents when stored at 2-8°C are stable up to expiry date printed on the package. Do not freeze.

Opened vials are stable for 30 days when stored at 2-8°C.

Concentrations in the test

rabbit brain phospholipid	1.2 %
magnesium aluminium silica	0.03 %
phenol	0.4 %
buffers	0.8 %

Warnings and notes

- Product for in vitro diagnostic use only.
- The reagents must be used only for the purpose intended by suitably qualified laboratory personnel, under appropriate laboratory conditions.
- APTT-P-REAGENTS are designed to work at 37°. Frequently check the temperature of all heating elements.
- Mix gently before use.
- Always follow instrument manufacturer's instructions for proper maintenance.

ADDITIONAL EQUIPMENT

- CALCIUM CHLORIDE reagent 0,025M (Cat. No K-355);
- a manual, mechanical or photo-optical means of clot detection;
- timer;
- control plasmas;
- general laboratory equipment

SPECIMEN

A. Anticoagulant- sodium citrate - 3.2% (0.105M).

B. Specimen Collection:

- Obtain venous blood.
- Immediately mix 9 parts blood with 1 part anticoagulant, mix well by inversion of tube.

3. Centrifuge the specimen at 1000 rcf for 15 min.

4. Remove plasma from the tube within 60 min using a plastic pipette and store in a plastic tube.

5. Test plasma sample within 2 hours, otherwise store frozen and thaw just prior to use.

Plasma pH will increase if exposed to air. Store samples stoppered. Do not delay mixing the blood with anticoagulant. Avoid foaming the specimen. Use only plastic containers.

PROCEDURE

Kit is suitable for use with manual, mechanical or automated instrument for clot detection. See instrument manufacturers instructions for full details.

Manual assay

- Prewarm CALCIUM CHLORIDE 0,025M to 37°C. Mix gently.
- Add 0.1 ml test plasma to cuvette and prewarm to 37°C.
- Add 0.1 ml APTT-EA-REAGENT to the test plasma. Mix and start timer.
- Incubate the plasma-reagent mixture at 37°C for 5 minutes (activation time). For consistent results, test all plasmas with the same activation time.
- Forcibly add 0.1 ml prewarmed CALCIUM CHLORIDE 0,025M.
- Note time for clot formation.
- Perform duplicate determinations.

Calculation

Calculate the mean clotting time of duplicate determinations (to the nearest 0.1 second) for each sample. A Normal Reference Range can also be reported for comparison. Do not report patient values relative to commercial control plasma clotting times. Controls are intended only for quality assurance of the test system, such as: temperature, reagents, pipettes, instrument etc.

REFERENCE VALUES

Method	Mean	Range (± SD)
mechanical	30.3	26.1-34.5
photo-optical	33.3	28.5-38.1

These values should only be used as a guideline. Each laboratory should establish a Normal Reference Range (NRR) using instrumentation, blood collection methods, and testing techniques used in that laboratory.

A new NRR should be established with any change in instrumentation, blood collection, techniques, anticoagulant and when changing to new lots of reagents.

QUALITY CONTROL

CONTROL PLASMA NORMAL (Cat. No K-100) CONTROL PLASMA-ABNORMAL LEVEL 1 and LEVEL 2 (Cat. No K-101 and K-102) should be tested in conjunction with patient plasmas. It is recommended that controls be run at least each shift and a minimum of once per 20 patient samples.

Each laboratory should establish a control range for each control.

LIMITATIONS

- Plasma samples with hematocrits outside the range of 20-55% may be improperly anticoagulated and should be adjusted appropriately.
- Turbid, icteric, lipemic, or hemolyzed specimens may generate erroneous results.
- Freezing and thawing plasma can affect results.
- Acute inflammatory reactions can shorten APTT results because of elevated fibrinogen.
- Sodium oxalate, EDTA, and heparin are not suitable anticoagulants. Oral contraceptives, estrogen, pregnancy, coumarin type drugs, heparin, asparaginase, and naloxone have been reported to influence APTT results.

PERFORMANCE CHARACTERISTICS

1. Heparin sensitivity:

For example, the following results were obtained on a photo-optical instrument with one lot of Cormay reagent:

Heparin conc. (U/ml)	APTT (s)
0.0	31.9
0.05	35.0
0.1	39.9
0.2	53.2
0.3	68.5
0.4	85.3
0.5	106.5

Each laboratory should establish its own heparin sensitivity ranges.

2. Factor sensitivity:

Cormay reagents was evaluated on mildly and severely deficient plasmas with the following results:

Factor	% activity	APTT (s)
VIII	<1%	150
VIII	20%	50.7
IX	<1%	>150
IX	20%	39.6
XI	<1%	88.7
XI	20%	44.5
XII	<1%	>150
XII	20%	43.2
Prekallikrein	<1%	55.9

Furthermore, the sensitivity of APTT-P to Factor VIII has been determined as follows:

% Factor VIII	APTT (s)
100%	34.3
70%	37.8
50%	40.9
40%	43.4
30%	46.2
20%	50.7
10%	57.0
5%	63.6
1%	85.2
<1%	>150

These values should only be used as guidelines. Each laboratory should establish sensitivity to individual factors using instruments, reagents, and techniques used in their laboratory.

WASTE MANAGEMENT

Please refer to local legal requirements.

LITERATURE

1. Langdell, R., Wagner, R., Brinkhous, K.: Effect of Antihe-mophilic Factor on One Stage Clotting Tests. J. Lab. Clin. Med. 41: 637, 1953.
2. Proctor, R., Rapaport, S.: The Partial Thromboplastin Time with Kaolin. Am. J. Clin. Path. 36: 212, 1961.
3. Ratnoff, O., Crum, S.: Activation of Hageman Factor by Solutions of Ellagic Acid. J. Lab. Clin. Med. 63: 359, 1964.
4. Brandt J.T Triplet! D.A.: Laboratory Monitoring of Heparin. Effect of Reagents and Instruments in the Activated Partial Thromboplastin Time. AmerJ Clin Path 76:530, 1981.
5. Thompson, J.M.: The Control of Heparin Therapy by the Activated Partial Thromboplastin Time. Sensitivity of Various Thromboplastins to Heparin. Standardization of Coagulation Assays: An Overview. Edited by D.A. Triplett, College of American Pathologists, Skokie, Ill. 1982, pp 195.
6. NCCLS: Collection, transport, and processing of blood specimens for coagulation testing and performance of coagulation assays. 3rd edition. Approved guideline. NCCLS Document H21-A3. Wayne, PA, 1998.
7. Young, D.S., Thomas, D.W., Friedman, R.B., et al: Effect of Drugs on Clinical Laboratory Tests. Clin Chem 18:1041, 1972.
8. Banez, E.L., Triplett, D.A., Koepke, J.: Laboratory Monitoring of Heparin Therapy The Effect of Different Salts of Heparin on the Activated Partial Thromboplastin Time. AmerJ Clin Path 74:569,1980.
9. Wujastyk, J., Triplett, D.A.: Selecting Instrumentation and Reagents for the Coagulation Laboratory. Pathologist 37:398, 1983.
10. Christensen, R.L., Triplett D.A.: Factor Assay (VIII and IX) Results in the College of American Pathologists Survey Program (1980-1982). AmerJ Clin Path 80 (Suppl): 633, 1983.
11. NCCLS: One-Stage Prothrombin time (PT) Test and Activated Partial Thromboplastin Time (APTT) Test; Approved Guideline. NCCLS Document H47-A. NCCLS, Wayne, PA, 1996.

Date of issue: 01. 2008.

MANUFACTURER

PZ CORMAY S.A.
ul. Wiosenna 22,
05-092 Łomianki, POLAND
tel.: +48 (0) 22 751 79 10
fax: +48 (0) 22 751 79 14
<http://www.pzcormay.pl>

01/08/01/08